


| Schedule to | |
|-------------------------------------|---|
| CERTIFICATE OF ACCREDITATION | |
| Laboratory | Eastern Bay Cylinders |
| Address | 7 Pohutu Street, Whakatane, 3120 |
| Telephone | 07 308-4077 |
| Fax | 07 308-4071 |
| Authorised Representative | Mr Ross Gildon Co-Managing Director |
| Client No. | 7171 |
| Programme | Gas Cylinder Testing Laboratory |
| Accreditation Number | 920 |
| Initial Accreditation Date | 15 October 2004 |
| Conformance Standard | NZS ISO/IEC 17025:2005 General requirements for the competence of testing and calibration laboratories |
| Testing Services Summary | 4.54 Cylinders and other Pressure Vessels |
| Signatories | Mr Ross Gildon 4.54 |

| | | | | |
|--------------------------------|---|---------|----------------|-------------|
| Authorised: General Manager |  | Issue 8 | Date: 22/07/13 | Page 1 of 2 |
|--------------------------------|---|---------|----------------|-------------|

Schedule to

CERTIFICATE OF ACCREDITATION

Eastern Bay Cylinders
 Gas Cylinder Testing Laboratory
SCOPE OF ACCREDITATION

Accreditation No 920

4.54 Cylinders and other Pressure Vessels

Tests on the following gas cylinders with water capacities from 2 litres to 200 litres:

Breathing Air

- SCUBA
- SCBA (excluding FRP)

Low Pressure Welded Cylinders

- Portable LPG - Externally valved cylinders (excluding stainless steel)
- Portable LPG - Internally valved cylinders ("Primus" type) – Visual only
- Refrigerant gases which have been purged prior to entry into the test station

Automotive

- LPG (excluding stainless steel)

Fire Extinguishers

- High Pressure (carbon dioxide)
- Low Pressure (thin walled)

The following tests in accordance with the Code of Practice for Periodic Testing of Compressed Gas Cylinders to AS 2337.1:2004 and AS 2337.3:2006


Mass check (4 kg to 250 kg)

- External examination
- Internal examination

Hydrostatic stretch and Proof pressure tests using the:

- (a) Non-water jacket method for pressures of 1.2 MPa to 6.0 MPa and 12 MPa to 60 MPa

Authorised:
 General Manager



Issue 8

Date: 22/07/13

Page 2 of 2